

CERTIFICATE OF COMPLIANCE

According to EMC 2004/108/EC & LVD 2006/95/EC

Applicant Name : Beijing Sevenstar Electronics Co.,Ltd.
Applicant Address : NO.1 Jiuxianqiao East Rd. Chaoyang District, Beijing
Manufacturer by : Beijing Sevenstar Electronics Co.,Ltd.
Manufacturer's Address : NO.1 Jiuxianqiao East Rd. Chaoyang District, Beijing
Brand Name : Sevenstar Electronics
Product Description : D08-1F,D08-1FM Flow Readout Box
Model / Part Number(s) : D08-1F,D08-1FM
Serial Number(s) Tested : Prototype
Test Report Number(s) : LEA-0806-10006-CE / LEA-0806-10006-SAFETY
Date(s) of Testing : 2008, June 28th to 29th /2008, July 15th to 18th

This is to certify that the product identified above is in compliance with the essential requirements of the following standards:

EN61326-1: 2006, Class B (Emission)
EN 61326-1: 2006, Conducted Emission.
EN 61326-1: 2006, Radiated Emission.
EN61000-3-2: 2006, Harmonic Current Emission.
EN61000-3-3: 1995 +A1:2001+A2:2005, Voltage Fluctuation
EN61326-1: 2006, (Immunity)
IEC 61000-4-2, Edition 1.2 (2001), Electrostatic Discharge Immunity
IEC 61000-4-3, Third Edition (2006), RF Immunity.
IEC 61000-4-4, Second Edition (2004), Electrical Fast Transient/Bursts Immunity Test.
IEC 61000-4-5, Second Edition (2005), Surge Immunity.
IEC 61000-4-6, Edition 2.2 (2006), Conducted Immunity.
IEC 61000-4-8, Edition 1.1(2001), Power Frequency Magnetic Field Immunity.
IEC 61000-4-11, Second Edition (2004), Voltage Dips and Short Interruptions Immunity.
EN 61010 - 1 : 2001, Safety requirements for electrical equipment for measurement, control, and laboratory use. Part 1: General requirements.

Reviewed by:

Signed by:

Paul J. Chen
Paul Chen / QC Manager

Issued by:



Issued Date: 2008, August 1st

This is the result of tests that were carried out from the submitted product sample(s) in conformity with the specification of the respective standards. The certificate holder has the right to affix the CE-Mark on the inspected product only when the product is completely complying with the required standards.

